

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051222\
 Data File : VW022993.D
 Acq On : 12 May 2022 11:18
 Operator : SY/VA
 Sample : N2766-12
 Misc : 6.89g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4G2

Quant Time: May 13 01:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 13 01:10:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	197202	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	107197	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	18769	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	79942	19.759	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.040%
7) Chloroethane-d5	2.885	69	62529	23.756	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.040%
11) 1,1-Dichloroethene-d2	4.013	63	67148	12.845	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.360%
21) 2-Butanone-d5	7.080	46	47405	61.625	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.240%
24) Chloroform-d	7.653	84	144437	25.000	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.000%
26) 1,2-Dichloroethane-d4	8.305	65	83052	26.210	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.840%
32) Benzene-d6	8.275	84	237298	38.405	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	153.640%#
36) 1,2-Dichloropropane-d6	9.275	67	81812	47.137	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	188.560%#
41) Toluene-d8	10.323	98	147865	24.773	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.080%
43) trans-1,3-Dichloroprop...	10.579	79	15472	19.360	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.440%
47) 2-Hexanone-d5	10.921	63	30688	92.621	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	185.240%#
56) 1,1,2,2-Tetrachloroeth...	12.689	84	69275	42.822	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	171.280%#
66) 1,2-Dichlorobenzene-d4	13.853	152	19728	30.069	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	120.280%#
Target Compounds						
13) Acetone	4.129	43	3217	5.586	ug/L	57
14) Carbon disulfide	4.385	76	31304	4.521	ug/L	97
42) Toluene	10.384	91	6728	0.951	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051222\
Data File : VW022993.D
Acq On : 12 May 2022 11:18
Operator : SY/VA
Sample : N2766-12
Misc : 6.89g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4G2

Quant Time: May 13 01:11:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 13 01:10:05 2022
Response via : Initial Calibration

