

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051222\
 Data File : VW022995.D
 Acq On : 12 May 2022 12:06
 Operator : SY/VA
 Sample : N2767-07
 Misc : 5.27g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGPM2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/15/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 13 01:11:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 13 01:10:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	372577	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	359983	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	172773	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	107959	14.123	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	56.480%		
7) Chloroethane-d5	2.885	69	84361	16.964	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	67.840%		
11) 1,1-Dichloroethene-d2	4.019	63	114943	11.638	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	46.560%		
21) 2-Butanone-d5	7.074	46	66821	45.977	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	91.960%		
24) Chloroform-d	7.647	84	225163	20.628	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	82.520%		
26) 1,2-Dichloroethane-d4	8.305	65	124589m	20.811	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	83.240%		
32) Benzene-d6	8.275	84	410085	19.764	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	79.040%		
36) 1,2-Dichloropropane-d6	9.274	67	125623	21.553	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	86.200%		
41) Toluene-d8	10.323	98	374068	18.663	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	74.640%		
43) trans-1,3-Dichloroprop...	10.579	79	49937	18.608	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	74.440%		
47) 2-Hexanone-d5	10.920	63	50174	45.094	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	90.180%		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	120464	22.174	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	88.680%		
66) 1,2-Dichlorobenzene-d4	13.853	152	133365	22.082	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	88.320%		
Target Compounds					Qvalue	
13) Acetone	4.129	43	4038	3.711	ug/L	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051222\
Data File : VW022995.D
Acq On : 12 May 2022 12:06
Operator : SY/VA
Sample : N2767-07
Misc : 5.27g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGPM2

Quant Time: May 13 01:11:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 13 01:10:05 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/15/2022
Supervised By : Mahesh Dadoda 05/16/2022

