

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051322\
 Data File : VW023022.D
 Acq On : 13 May 2022 13:08
 Operator : SY/VA
 Sample : N2821-03
 Misc : 4.77g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXLC2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/15/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 14 00:00:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 13 23:57:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	364016	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	288012	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	93304	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	200913	26.902	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 107.600%			
7) Chloroethane-d5	2.885	69	145919	30.032	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 120.120%			
11) 1,1-Dichloroethene-d2	4.019	63	174866m	18.121	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 72.480%			
21) 2-Butanone-d5	7.080	46	38933	27.418	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 54.840%			
24) Chloroform-d	7.647	84	292623	27.438	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 109.760%			
26) 1,2-Dichloroethane-d4	8.305	65	169713	29.015	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 116.040%			
32) Benzene-d6	8.275	84	559462	33.701	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 134.800%			
36) 1,2-Dichloropropane-d6	9.274	67	169835	36.420	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 145.680%#			
41) Toluene-d8	10.323	98	471608	29.408	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 117.640%			
43) trans-1,3-Dichloroprop...	10.579	79	43029	20.040	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 80.160%			
47) 2-Hexanone-d5	10.927	63	27803	31.232	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 62.460%			
56) 1,1,2,2-Tetrachloroeth...	12.688	84	123432	28.398	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 113.600%			
66) 1,2-Dichlorobenzene-d4	13.847	152	90837	27.850	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 111.400%			
Target Compounds					Qvalue	
13) Acetone	4.129	43	4027	3.788	ug/L	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051322\
Data File : VW023022.D
Acq On : 13 May 2022 13:08
Operator : SY/VA
Sample : N2821-03
Misc : 4.77g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXLC2

Quant Time: May 14 00:00:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 13 23:57:51 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/15/2022
Supervised By :Mahesh Dadoda 05/16/2022

