

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051322\
 Data File : VW023025.D
 Acq On : 13 May 2022 14:21
 Operator : SY/VA
 Sample : N2821-06
 Misc : 4.66g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXLC6

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/15/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 14 00:00:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 13 23:57:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	380707	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	316790	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	109463	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	185958	23.808	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 95.240%			
7) Chloroethane-d5	2.885	69	132558	26.086	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 104.360%			
11) 1,1-Dichloroethene-d2	4.013	63	165712	16.420	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 65.680%			
21) 2-Butanone-d5	7.080	46	35127	23.653	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 47.300%			
24) Chloroform-d	7.647	84	268456	24.069	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 96.280%			
26) 1,2-Dichloroethane-d4	8.305	65	157258m	25.707	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 102.840%			
32) Benzene-d6	8.275	84	513616	28.129	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 112.520%			
36) 1,2-Dichloropropane-d6	9.275	67	152862	29.803	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 119.200%			
41) Toluene-d8	10.323	98	451941	25.622	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 102.480%			
43) trans-1,3-Dichloroprop...	10.579	79	49706	21.047	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 84.200%			
47) 2-Hexanone-d5	10.921	63	25065	25.599	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 51.200%			
56) 1,1,2,2-Tetrachloroeth...	12.689	84	119598	25.017	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 100.080%			
66) 1,2-Dichlorobenzene-d4	13.847	152	97049	25.363	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 101.440%			
Target Compounds					Qvalue	
13) Acetone	4.129	43	3514	3.160	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051322\
Data File : VW023025.D
Acq On : 13 May 2022 14:21
Operator : SY/VA
Sample : N2821-06
Misc : 4.66g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXLC6

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/15/2022
Supervised By : Mahesh Dadoda 05/16/2022

Quant Time: May 14 00:00:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 13 23:57:51 2022
Response via : Initial Calibration

