

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051722\
 Data File : VW023089.D
 Acq On : 17 May 2022 10:35
 Operator : SY/VA
 Sample : N2843-17
 Misc : 6.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4J7

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/18/2022
 Supervised By :Mahesh Dadoda 05/26/2022

Quant Time: May 18 01:27:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 18 01:26:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	278417	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	253631	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	105804	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	78001	13.655	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	54.640%		
7) Chloroethane-d5	2.885	69	62741	16.883	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	67.520%		
11) 1,1-Dichloroethene-d2	4.019	63	86930	11.778	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	47.120%		
21) 2-Butanone-d5	7.080	46	57186	52.655	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	105.300%		
24) Chloroform-d	7.653	84	163409	20.033	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.120%		
26) 1,2-Dichloroethane-d4	8.305	65	97112m	21.707	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	86.840%		
32) Benzene-d6	8.275	84	290817	19.893	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	79.560%		
36) 1,2-Dichloropropane-d6	9.275	67	92994	22.645	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	90.600%		
41) Toluene-d8	10.323	98	260664	18.458	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	73.840%		
43) trans-1,3-Dichloroprop...	10.579	79	37020	19.579	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	78.320%		
47) 2-Hexanone-d5	10.927	63	45334	57.829	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	115.660%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	94219	24.616	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	98.480%		
66) 1,2-Dichlorobenzene-d4	13.853	152	73151	19.778	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	79.120%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051722\
Data File : VW023089.D
Acq On : 17 May 2022 10:35
Operator : SY/VA
Sample : N2843-17
Misc : 6.10g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4J7

Quant Time: May 18 01:27:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 18 01:26:16 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/18/2022
Supervised By : Mahesh Dadoda 05/26/2022

