

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051722\
 Data File : VW023088.D
 Acq On : 17 May 2022 10:10
 Operator : SY/VA
 Sample : N2821-23
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: May 18 01:27:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 18 01:26:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	343084	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	333422	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	168709	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	154797	21.992	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.960%
7) Chloroethane-d5	2.885	69	101137	22.085	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.360%
11) 1,1-Dichloroethene-d2	4.019	63	156464	17.203	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.800%
21) 2-Butanone-d5	7.080	46	60296	45.054	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.100%
24) Chloroform-d	7.653	84	246705	24.544	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.160%
26) 1,2-Dichloroethane-d4	8.305	65	135428	24.566	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.280%
32) Benzene-d6	8.275	84	487967	25.391	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.560%
36) 1,2-Dichloropropane-d6	9.275	67	137812	25.528	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.120%
41) Toluene-d8	10.323	98	466685	25.138	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.560%
43) trans-1,3-Dichloroprop...	10.579	79	60076	24.169	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.680%
47) 2-Hexanone-d5	10.921	63	48797	47.350	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.700%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	119309	23.711	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	157859	26.767	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051722\
Data File : VW023088.D
Acq On : 17 May 2022 10:10
Operator : SY/VA
Sample : N2821-23
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: May 18 01:27:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 18 01:26:16 2022
Response via : Initial Calibration

