

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051822\
 Data File : VW023131.D
 Acq On : 18 May 2022 14:58
 Operator : SY/VA
 Sample : N2860-07
 Misc : 6.89g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 H0AT3

Quant Time: May 19 01:24:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu May 19 01:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	341439	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	326579	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	169817	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	85294	12.176	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	48.720%
7) Chloroethane-d5	2.885	69	64348	14.119	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.480%
11) 1,1-Dichloroethene-d2	4.013	63	95011	10.497	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	42.000%#
21) 2-Butanone-d5	7.080	46	63273	47.506	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.020%
24) Chloroform-d	7.647	84	191665	19.160	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.640%
26) 1,2-Dichloroethane-d4	8.305	65	107440	19.583	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.320%
32) Benzene-d6	8.275	84	342600	18.200	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	72.800%
36) 1,2-Dichloropropane-d6	9.275	67	105142	19.884	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.520%
41) Toluene-d8	10.323	98	310012	17.049	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.200%
43) trans-1,3-Dichloroprop...	10.579	79	39722	16.315	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.280%
47) 2-Hexanone-d5	10.921	63	49724	49.261	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.520%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	116872	23.714	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	113677	19.150	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051822\
Data File : VW023131.D
Acq On : 18 May 2022 14:58
Operator : SY/VA
Sample : N2860-07
Misc : 6.89g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
H0AT3

Quant Time: May 19 01:24:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu May 19 01:20:33 2022
Response via : Initial Calibration

