

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051922\
 Data File : VW023174.D
 Acq On : 19 May 2022 15:43
 Operator : SY/VA
 Sample : N2925-19
 Misc : 4.74g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXXZ6

Quant Time: May 20 01:51:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 20 01:47:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	222761	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	198251	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	100184	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	63067	13.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.200%
7) Chloroethane-d5	2.885	69	59041	19.857	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.440%
11) 1,1-Dichloroethene-d2	4.013	63	74273	12.577	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	50.320%
21) 2-Butanone-d5	7.073	46	72621	83.573	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	167.140%#
24) Chloroform-d	7.653	84	189795	29.081	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	116.320%
26) 1,2-Dichloroethane-d4	8.305	65	111904	31.263	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	125.040%
32) Benzene-d6	8.274	84	327139	28.628	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	114.520%
36) 1,2-Dichloropropane-d6	9.274	67	111077	34.605	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	138.400%#
41) Toluene-d8	10.323	98	270992	24.550	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.200%
43) trans-1,3-Dichloroprop...	10.579	79	39114	26.465	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	105.840%
47) 2-Hexanone-d5	10.926	63	51889	84.680	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	169.360%#
56) 1,1,2,2-Tetrachloroeth...	12.694	84	114582	38.298	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	153.200%#
66) 1,2-Dichlorobenzene-d4	13.853	152	107264	30.629	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	122.520%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051922\
Data File : VW023174.D
Acq On : 19 May 2022 15:43
Operator : SY/VA
Sample : N2925-19
Misc : 4.74g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXXZ6

Quant Time: May 20 01:51:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 20 01:47:53 2022
Response via : Initial Calibration

