

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051922\
 Data File : VW023178.D
 Acq On : 19 May 2022 17:19
 Operator : SY/VA
 Sample : N2848-21RE
 Misc : 6.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXY10

Quant Time: May 20 01:52:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 20 01:47:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	406447	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	345424	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	126303	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	174339	20.907	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.640%
7) Chloroethane-d5	2.891	69	121409	22.379	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.520%
11) 1,1-Dichloroethene-d2	4.019	63	167342	15.531	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.120%
21) 2-Butanone-d5	7.074	46	61071	38.519	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.040%
24) Chloroform-d	7.653	84	259855	21.822	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.280%
26) 1,2-Dichloroethane-d4	8.305	65	146969	22.503	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.000%
32) Benzene-d6	8.275	84	521310	26.183	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.720%
36) 1,2-Dichloropropane-d6	9.274	67	149187	26.675	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.680%
41) Toluene-d8	10.323	98	471638	24.522	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.080%
43) trans-1,3-Dichloroprop...	10.579	79	57490	22.325	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.280%
47) 2-Hexanone-d5	10.921	63	50289	47.102	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.200%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	111494	21.388	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.560%
66) 1,2-Dichlorobenzene-d4	13.853	152	99304	22.492	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051922\
Data File : VW023178.D
Acq On : 19 May 2022 17:19
Operator : SY/VA
Sample : N2848-21RE
Misc : 6.03g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXY10

Quant Time: May 20 01:52:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 20 01:47:53 2022
Response via : Initial Calibration

