

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051923\
 Data File : VW026001.D
 Acq On : 19 May 2023 13:24
 Operator : SY/MD
 Sample : 02859-02
 Misc : 6.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C11G0

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/22/2023
 Supervised By :Mahesh Dadoda 05/22/2023

Quant Time: May 20 05:07:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 20 04:57:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	685615	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	642492	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	284158	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	189034	19.457	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	77.840%		
7) Chloroethane-d5	2.899	69	159867	22.195	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	88.760%		
11) 1,1-Dichloroethene-d2	4.021	65	93860	19.293	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	77.160%		
21) 2-Butanone-d5	7.075	46	163646	50.950	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	101.900%		
24) Chloroform-d	7.648	84	434886m	21.997	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	88.000%		
26) 1,2-Dichloroethane-d4	8.306	65	253809	22.703	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	90.800%		
32) Benzene-d6	8.276	84	855670	20.382	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	81.520%		
36) 1,2-Dichloropropane-d6	9.276	67	291533	21.704	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	86.800%		
41) Toluene-d8	10.324	98	727590	19.436	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	77.760%		
43) trans-1,3-Dichloroprop...	10.574	79	107840	19.806	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	79.240%		
47) 2-Hexanone-d5	10.922	63	108955	47.569	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	95.140%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	247817	24.206	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	96.840%		
66) 1,2-Dichlorobenzene-d4	13.848	152	253049	24.429	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	97.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051923\
Data File : VW026001.D
Acq On : 19 May 2023 13:24
Operator : SY/MD
Sample : 02859-02
Misc : 6.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C11G0

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/22/2023
Supervised By :Mahesh Dadoda 05/22/2023

Quant Time: May 20 05:07:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 20 04:57:36 2023
Response via : Initial Calibration

