

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
 Data File : VW023205.D
 Acq On : 20 May 2022 17:46
 Operator : SY/VA
 Sample : N2925-11
 Misc : 4.74g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXXY8

Quant Time: May 20 23:06:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 20 23:01:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	378368	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	343692	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	127386	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	156017	20.098	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.400%
7) Chloroethane-d5	2.885	69	104731	20.738	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.960%
11) 1,1-Dichloroethene-d2	4.019	63	140392	13.997	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.000%
21) 2-Butanone-d5	7.074	46	59981	40.639	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.280%
24) Chloroform-d	7.647	84	258145	23.287	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.160%
26) 1,2-Dichloroethane-d4	8.384	65	504	0.083	ug/L	0.08
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.320%#
32) Benzene-d6	8.275	84	473263	23.890	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.560%
36) 1,2-Dichloropropane-d6	9.274	67	147056	26.426	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.720%
41) Toluene-d8	10.323	98	416024	21.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.960%
43) trans-1,3-Dichloroprop...	10.579	79	51382	20.054	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.200%
47) 2-Hexanone-d5	10.921	63	47007	44.250	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.500%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	122417	23.602	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	116497	26.162	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
Data File : VW023205.D
Acq On : 20 May 2022 17:46
Operator : SY/VA
Sample : N2925-11
Misc : 4.74g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXXY8

Quant Time: May 20 23:06:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 20 23:01:40 2022
Response via : Initial Calibration

