

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
 Data File : VW023217.D
 Acq On : 20 May 2022 23:24
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK471

Quant Time: May 21 01:50:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 21 01:30:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	291929	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	297713	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	147630	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	126470	21.116	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.480%
7) Chloroethane-d5	2.879	69	68756	17.645	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.600%
11) 1,1-Dichloroethene-d2	4.013	63	118890	15.363	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.440%
21) 2-Butanone-d5	7.074	46	53759	47.208	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.420%
24) Chloroform-d	7.647	84	209193	24.459	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.840%
26) 1,2-Dichloroethane-d4	8.305	65	119481	25.471	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.880%
32) Benzene-d6	8.275	84	408915	23.830	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.320%
36) 1,2-Dichloropropane-d6	9.274	67	118302	24.543	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.160%
41) Toluene-d8	10.323	98	387460	23.374	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.480%
43) trans-1,3-Dichloroprop...	10.579	79	48514	21.859	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.440%
47) 2-Hexanone-d5	10.927	63	42550	46.241	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.480%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	106544	23.714	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	132415	25.659	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
Data File : VW023217.D
Acq On : 20 May 2022 23:24
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK471

Quant Time: May 21 01:50:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 21 01:30:17 2022
Response via : Initial Calibration

