

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
 Data File : VW023224.D
 Acq On : 21 May 2022 02:12
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK473

Quant Time: May 21 04:07:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 21 01:30:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	301686	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	302312	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	144414	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	127128	20.539	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.160%
7) Chloroethane-d5	2.879	69	79474	19.736	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.960%
11) 1,1-Dichloroethene-d2	4.019	63	119456	14.937	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.760%
21) 2-Butanone-d5	7.079	46	55895	47.496	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.000%
24) Chloroform-d	7.646	84	207218	23.444	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.760%
26) 1,2-Dichloroethane-d4	8.305	65	120601	24.878	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.520%
32) Benzene-d6	8.274	84	413487	23.729	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.920%
36) 1,2-Dichloropropane-d6	9.274	67	119607	24.436	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.760%
41) Toluene-d8	10.323	98	391517	23.259	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.040%
43) trans-1,3-Dichloroprop...	10.579	79	48232	21.401	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.600%
47) 2-Hexanone-d5	10.926	63	43650	46.714	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.420%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	108192	23.715	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	133554	26.456	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
Data File : VW023224.D
Acq On : 21 May 2022 02:12
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK473

Quant Time: May 21 04:07:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 21 01:30:17 2022
Response via : Initial Calibration

