

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
 Data File : VW023189.D
 Acq On : 20 May 2022 11:20
 Operator : SY/VA
 Sample : N2883-10
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: May 20 23:02:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 20 23:01:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	347419	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	330171	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	172627	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	148666	20.857	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.440%
7) Chloroethane-d5	2.885	69	88460	19.076	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.320%
11) 1,1-Dichloroethene-d2	4.013	63	159210	17.287	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.160%
21) 2-Butanone-d5	7.080	46	59061	43.580	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.160%
24) Chloroform-d	7.647	84	242154	23.791	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.160%
26) 1,2-Dichloroethane-d4	8.305	65	134328	24.062	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.240%
32) Benzene-d6	8.274	84	492698	25.889	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.560%
36) 1,2-Dichloropropane-d6	9.274	67	138774	25.959	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.840%
41) Toluene-d8	10.323	98	471246	25.634	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.520%
43) trans-1,3-Dichloroprop...	10.579	79	56493	22.951	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.800%
47) 2-Hexanone-d5	10.920	63	44407	43.515	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.020%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	116104	23.302	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	157409	26.085	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
Data File : VW023189.D
Acq On : 20 May 2022 11:20
Operator : SY/VA
Sample : N2883-10
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: May 20 23:02:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 20 23:01:40 2022
Response via : Initial Calibration

