

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
 Data File : VW023194.D
 Acq On : 20 May 2022 13:21
 Operator : SY/VA
 Sample : VIBLK469
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK469

Quant Time: May 20 23:04:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 20 23:01:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	338457	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	339341	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	171137	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	136002	19.586	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.360%
7) Chloroethane-d5	2.885	69	84805	18.772	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.080%
11) 1,1-Dichloroethene-d2	4.019	63	139914	15.594	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.360%
21) 2-Butanone-d5	7.079	46	59352	44.955	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.900%
24) Chloroform-d	7.652	84	227572	22.950	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.800%
26) 1,2-Dichloroethane-d4	8.305	65	126678	23.293	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.160%
32) Benzene-d6	8.274	84	454753	23.250	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	9.274	67	129589	23.586	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.360%
41) Toluene-d8	10.323	98	445541	23.581	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.320%
43) trans-1,3-Dichloroprop...	10.579	79	53827	21.277	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.120%
47) 2-Hexanone-d5	10.926	63	47931	45.699	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.400%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	113529	22.169	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	148109	24.758	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.040%
Target Compounds						Qvalue
40) 4-Methyl-2-pentanone	10.213	43	5543	1.818	ug/L	100
48) 2-Hexanone	10.969	43	5282	2.593	ug/L #	94
72) 1,2,3-Trichlorobenzene	15.547	180	3922	0.792	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
Data File : VW023194.D
Acq On : 20 May 2022 13:21
Operator : SY/VA
Sample : VIBLK469
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK469

Quant Time: May 20 23:04:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 20 23:01:40 2022
Response via : Initial Calibration

