

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
 Data File : VW023218.D
 Acq On : 20 May 2022 23:48
 Operator : SY/VA
 Sample : N2925-20
 Misc : 4.69g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXXZ7

Quant Time: May 21 01:51:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 21 01:30:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	352749	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	349851	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	168932	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	131662	18.192	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.760%
7) Chloroethane-d5	2.885	69	82677	17.560	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.240%
11) 1,1-Dichloroethene-d2	4.019	63	131435	14.055	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.240%
21) 2-Butanone-d5	7.074	46	65586	47.664	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.320%
24) Chloroform-d	7.647	84	248186	24.015	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.040%
26) 1,2-Dichloroethane-d4	8.305	65	143745	25.360	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.440%
32) Benzene-d6	8.275	84	472815	23.447	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.800%
36) 1,2-Dichloropropane-d6	9.274	67	142616	25.177	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.720%
41) Toluene-d8	10.323	98	433662	22.262	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.040%
43) trans-1,3-Dichloroprop...	10.579	79	55708	21.359	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.440%
47) 2-Hexanone-d5	10.920	63	51986	48.076	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.160%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	128871	24.409	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.640%
66) 1,2-Dichlorobenzene-d4	13.853	152	152275	25.786	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052022\
Data File : VW023218.D
Acq On : 20 May 2022 23:48
Operator : SY/VA
Sample : N2925-20
Misc : 4.69g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXXZ7

Quant Time: May 21 01:51:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 21 01:30:17 2022
Response via : Initial Calibration

