

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052322\
 Data File : VW023255.D
 Acq On : 23 May 2022 17:12
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK476

Quant Time: May 24 01:31:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 24 01:27:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	275432	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	278167	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	139131	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	123799	21.908	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.640%
7) Chloroethane-d5	2.879	69	80763	21.968	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.880%
11) 1,1-Dichloroethene-d2	4.013	63	109679	15.021	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.080%
21) 2-Butanone-d5	7.080	46	55613	51.761	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.520%
24) Chloroform-d	7.647	84	189632	23.500	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	8.305	65	109782	24.805	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.200%
32) Benzene-d6	8.275	84	366691	22.871	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.480%
36) 1,2-Dichloropropane-d6	9.275	67	108258	24.037	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.160%
41) Toluene-d8	10.323	98	351434	22.690	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.760%
43) trans-1,3-Dichloroprop...	10.573	79	44683	21.547	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.200%
47) 2-Hexanone-d5	10.927	63	44331	51.561	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.120%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	103360	24.622	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.480%
66) 1,2-Dichlorobenzene-d4	13.853	152	120801	24.838	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052322\
Data File : VW023255.D
Acq On : 23 May 2022 17:12
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK476

Quant Time: May 24 01:31:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 24 01:27:20 2022
Response via : Initial Calibration

