

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052322\
 Data File : VW023257.D
 Acq On : 23 May 2022 18:13
 Operator : SY/VA
 Sample : VIBLK477
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK477

Quant Time: May 24 01:32:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 24 01:27:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	287317	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	293118	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	147541	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	126955	21.537	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.160%
7) Chloroethane-d5	2.879	69	76792	20.024	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.080%
11) 1,1-Dichloroethene-d2	4.013	63	104722	13.749	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.000%
21) 2-Butanone-d5	7.074	46	49863	44.490	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.980%
24) Chloroform-d	7.647	84	189764	22.543	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.160%
26) 1,2-Dichloroethane-d4	8.305	65	112339	24.333	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.320%
32) Benzene-d6	8.275	84	369427	21.866	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.480%
36) 1,2-Dichloropropane-d6	9.274	67	107084	22.564	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.240%
41) Toluene-d8	10.323	98	348813	21.372	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.480%
43) trans-1,3-Dichloroprop...	10.579	79	43709	20.002	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.000%
47) 2-Hexanone-d5	10.921	63	40736	44.963	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.920%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	100864	22.802	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	116246	22.539	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052322\
Data File : VW023257.D
Acq On : 23 May 2022 18:13
Operator : SY/VA
Sample : VIBLK477
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK477

Quant Time: May 24 01:32:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 24 01:27:20 2022
Response via : Initial Calibration

