

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052422\
 Data File : VW023292.D
 Acq On : 25 May 2022 00:26
 Operator : SY/VA
 Sample : N2995-05
 Misc : 5.22g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBGQ8

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/26/2022
 Supervised By :Mahesh Dadoda 05/26/2022

Quant Time: May 25 05:58:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 25 04:05:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	311885	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	278628	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	111283	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	127739	19.963	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 79.840%			
7) Chloroethane-d5	2.879	69	80243	19.276	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 77.120%			
11) 1,1-Dichloroethene-d2	4.013	63	120714	14.600	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 58.400%			
21) 2-Butanone-d5	7.080	46	98377	80.861	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 161.720%#			
24) Chloroform-d	7.647	84	259320	28.380	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 113.520%			
26) 1,2-Dichloroethane-d4	8.305	65	165584m	33.040	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 132.160%#			
32) Benzene-d6	8.275	84	439797	27.385	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 109.520%			
36) 1,2-Dichloropropane-d6	9.274	67	151980	33.689	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 134.760%#			
41) Toluene-d8	10.323	98	344385	22.198	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 88.800%			
43) trans-1,3-Dichloroprop...	10.579	79	54560	26.266	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 105.080%			
47) 2-Hexanone-d5	10.921	63	73318	85.135	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 170.280%#			
56) 1,1,2,2-Tetrachloroeth...	12.688	84	172981	41.139	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 164.560%#			
66) 1,2-Dichlorobenzene-d4	13.853	152	116280	29.891	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 119.560%			
Target Compounds					Qvalue	
13) Acetone	4.129	43	9887	10.855	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052422\
Data File : VW023292.D
Acq On : 25 May 2022 00:26
Operator : SY/VA
Sample : N2995-05
Misc : 5.22g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GBGQ8

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/26/2022
Supervised By : Mahesh Dadoda 05/26/2022

Quant Time: May 25 05:58:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 25 04:05:49 2022
Response via : Initial Calibration

