

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052422\
 Data File : VW023296.D
 Acq On : 25 May 2022 02:02
 Operator : SY/VA
 Sample : N2995-09
 Misc : 3.70g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBGR2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/26/2022
 Supervised By : Mahesh Dadoda 05/26/2022

Quant Time: May 25 05:59:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 25 04:05:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	337221	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	308614	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	126419	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	125030	18.072	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.280%
7) Chloroethane-d5	2.879	69	71433	15.870	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.480%
11) 1,1-Dichloroethene-d2	4.013	63	119459	13.363	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.440%
21) 2-Butanone-d5	7.080	46	90129	68.516	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	137.040%#
24) Chloroform-d	7.647	84	250731	25.378	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	101.520%
26) 1,2-Dichloroethane-d4	8.305	65	155535m	28.703	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	114.800%
32) Benzene-d6	8.275	84	428897	24.111	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.440%
36) 1,2-Dichloropropane-d6	9.274	67	146571	29.333	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	117.320%
41) Toluene-d8	10.323	98	346845	20.185	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.720%
43) trans-1,3-Dichloroprop...	10.579	79	52895	22.991	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.960%
47) 2-Hexanone-d5	10.920	63	69625	72.992	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	145.980%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	159049	34.150	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	136.600%#
66) 1,2-Dichlorobenzene-d4	13.847	152	118469	26.808	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.240%
Target Compounds					Qvalue	
13) Acetone	4.123	43	32136	32.630	ug/L	97
34) Trichloroethene	9.098	95	3231	0.646	ug/L #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052422\
Data File : VW023296.D
Acq On : 25 May 2022 02:02
Operator : SY/VA
Sample : N2995-09
Misc : 3.70g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GBGR2

Quant Time: May 25 05:59:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 25 04:05:49 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/26/2022
Supervised By : Mahesh Dadoda 05/26/2022

