

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052523\
 Data File : VW026041.D
 Acq On : 25 May 2023 15:26
 Operator : SY/MD
 Sample : 02950-04
 Misc : 3.18g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 EW981

Quant Time: May 26 00:34:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 26 00:33:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	814000	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	636582	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	215881	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	220772	19.139	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	76.560%		
7) Chloroethane-d5	2.893	69	207526	24.267	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	97.080%		
11) 1,1-Dichloroethene-d2	4.015	65	116992	20.255	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	81.040%		
21) 2-Butanone-d5	7.075	46	211844	55.554	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	111.100%		
24) Chloroform-d	7.648	84	561390	23.918	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	95.680%		
26) 1,2-Dichloroethane-d4	8.307	65	319180	24.048	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	96.200%		
32) Benzene-d6	8.276	84	1096382	26.358	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	105.440%		
36) 1,2-Dichloropropane-d6	9.276	67	373034	28.029	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	112.120%		
41) Toluene-d8	10.325	98	877427	23.656	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	94.640%		
43) trans-1,3-Dichloroprop...	10.575	79	118948	22.049	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	88.200%		
47) 2-Hexanone-d5	10.922	63	141615	62.402	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	124.800%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	266861	26.308	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	105.240%		
66) 1,2-Dichlorobenzene-d4	13.849	152	194585	24.726	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	98.920%		
Target Compounds						
13) Acetone	4.125	43	53198	15.733	ug/L	100
16) Methylene chloride	4.917	84	107356	6.395	ug/L	93
22) 2-Butanone	7.173	43	20372	4.363	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052523\
 Data File : VW026041.D
 Acq On : 25 May 2023 15:26
 Operator : SY/MD
 Sample : 02950-04
 Misc : 3.18g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW981

Quant Time: May 26 00:34:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 26 00:33:05 2023
 Response via : Initial Calibration

