

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052924\
 Data File : VW028875.D
 Acq On : 29 May 2024 11:39
 Operator : SY/MD
 Sample : P2591-01
 Misc : 5.13g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH5W0

Quant Time: May 30 01:30:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 29 23:53:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	441778	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	343744	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	93617	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	101091	12.050	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	48.200%		
7) Chloroethane-d5	2.899	69	98894	15.396	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	61.600%		
11) 1,1-Dichloroethene-d2	4.021	65	49462	14.112	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	56.440%		
21) 2-Butanone-d5	7.081	46	78666	35.943	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	71.880%		
24) Chloroform-d	7.648	84	280328	20.308	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	81.240%		
26) 1,2-Dichloroethane-d4	8.307	65	150785	19.080	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	76.320%		
32) Benzene-d6	8.276	84	497992	21.714	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	86.840%		
36) 1,2-Dichloropropane-d6	9.270	67	172115	24.940	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	99.760%		
41) Toluene-d8	10.325	98	395121	19.454	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	77.800%		
43) trans-1,3-Dichloroprop...	10.581	79	48129	17.336	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	69.360%		
47) 2-Hexanone-d5	10.922	63	48374	40.641	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	81.280%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	118381	20.457	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	81.840%		
66) 1,2-Dichlorobenzene-d4	13.848	152	80119	22.454	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	89.800%		
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	30254	4.428	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052924\
Data File : VW028875.D
Acq On : 29 May 2024 11:39
Operator : SY/MD
Sample : P2591-01
Misc : 5.13g/10mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH5W0

Quant Time: May 30 01:30:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 29 23:53:47 2024
Response via : Initial Calibration

