

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052924\
 Data File : VW028897.D
 Acq On : 29 May 2024 21:20
 Operator : SY/MD
 Sample : P2647-06
 Misc : 6.37g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH5T7

Quant Time: May 30 01:34:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 29 23:53:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	443042	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	413611	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	159978	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	136884	16.270	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.080%
7) Chloroethane-d5	2.905	69	122826	19.067	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.280%
11) 1,1-Dichloroethene-d2	4.027	65	65489	18.631	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.520%
21) 2-Butanone-d5	7.075	46	82575	37.621	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.240%
24) Chloroform-d	7.648	84	316795	22.884	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.520%
26) 1,2-Dichloroethane-d4	8.307	65	174864	22.063	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.240%
32) Benzene-d6	8.276	84	617260	22.368	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.480%
36) 1,2-Dichloropropane-d6	9.276	67	196522	23.666	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.680%
41) Toluene-d8	10.325	98	528544	21.628	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.520%
43) trans-1,3-Dichloroprop...	10.575	79	64797	19.397	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.600%
47) 2-Hexanone-d5	10.922	63	56691	39.583	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.160%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	140277	20.146	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	13.848	152	146663	24.053	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.200%
Target Compounds						
16) Methylene chloride	4.917	84	17303	2.525	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052924\
Data File : VW028897.D
Acq On : 29 May 2024 21:20
Operator : SY/MD
Sample : P2647-06
Misc : 6.37g/10mL/MSVOA_W/SOIL/A
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH5T7

Quant Time: May 30 01:34:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 29 23:53:47 2024
Response via : Initial Calibration

