

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060222\
 Data File : VW023374.D
 Acq On : 02 Jun 2022 15:04
 Operator : SY/VA
 Sample : VW0602SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK521

Quant Time: Jun 02 21:57:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 02 21:56:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	393538	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	367542	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	175314	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	138956	23.876	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.520%
7) Chloroethane-d5	2.885	69	101814	25.163	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.640%
11) 1,1-Dichloroethene-d2	4.153	63	335	0.037	ug/L	0.13
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.160%#
21) 2-Butanone-d5	7.074	46	56191	46.909	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.820%
24) Chloroform-d	7.653	84	222497	21.470	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.880%
26) 1,2-Dichloroethane-d4	8.305	65	124831	22.998	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.000%
32) Benzene-d6	8.275	84	439929	22.277	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.120%
36) 1,2-Dichloropropane-d6	9.274	67	128489	22.742	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.960%
41) Toluene-d8	10.323	98	417712	22.184	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.720%
43) trans-1,3-Dichloroprop...	10.579	79	53134	21.067	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.280%
47) 2-Hexanone-d5	10.927	63	41741	45.988	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.980%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	108802	21.834	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.320%
66) 1,2-Dichlorobenzene-d4	13.853	152	142345	22.970	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060222\
Data File : VW023374.D
Acq On : 02 Jun 2022 15:04
Operator : SY/VA
Sample : VW0602SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK521

Quant Time: Jun 02 21:57:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 02 21:56:41 2022
Response via : Initial Calibration

