

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060324\
 Data File : VW028971.D
 Acq On : 03 Jun 2024 18:18
 Operator : SY/MD
 Sample : P2678-14
 Misc : 5.35g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH683

Quant Time: Jun 04 02:21:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 04 02:14:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	433930	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	308626	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	77363	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	178744	25.929	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 103.720%	
7) Chloroethane-d5	2.899	69	141422	26.716	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.880%	
11) 1,1-Dichloroethene-d2	4.021	65	73642	24.305	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 97.240%	
21) 2-Butanone-d5	7.075	46	72551	51.511	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 103.020%	
24) Chloroform-d	7.648	84	319226	25.786	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 103.160%	
26) 1,2-Dichloroethane-d4	8.307	65	171170	26.263	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 105.040%	
32) Benzene-d6	8.276	84	630601	33.290	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 133.160%	
36) 1,2-Dichloropropane-d6	9.276	67	203490	35.430	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 141.720%#	
41) Toluene-d8	10.319	98	497076	29.406	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 117.640%	
43) trans-1,3-Dichloroprop...	10.575	79	51301	23.245	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 93.000%	
47) 2-Hexanone-d5	10.922	63	46242	61.012	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 122.020%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	126733	31.763	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 127.040%#	
66) 1,2-Dichlorobenzene-d4	13.848	152	77555	28.250	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 113.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060324\
Data File : VW028971.D
Acq On : 03 Jun 2024 18:18
Operator : SY/MD
Sample : P2678-14
Misc : 5.35g/10mL/MSVOA_W/SOIL/A
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH683

Quant Time: Jun 04 02:21:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 04 02:14:06 2024
Response via : Initial Calibration

