

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060324\
 Data File : VW028976.D
 Acq On : 03 Jun 2024 20:18
 Operator : SY/MD
 Sample : P2678-19
 Misc : 5.48g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH688

Quant Time: Jun 04 02:22:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 04 02:14:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	145754	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	110826	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	28407	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	69635	30.073	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	= 120.280%	
7) Chloroethane-d5	2.893	69	56132	31.569	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 126.280%	
11) 1,1-Dichloroethene-d2	4.015	65	27719	27.236	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 108.960%	
21) 2-Butanone-d5	7.081	46	24240	51.237	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 102.480%	
24) Chloroform-d	7.648	84	123543	29.710	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 118.840%	
26) 1,2-Dichloroethane-d4	8.307	65	71090	32.473	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 129.880%	
32) Benzene-d6	8.276	84	233210	34.284	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 137.120%#	
36) 1,2-Dichloropropane-d6	9.270	67	78956	38.283	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 153.120%#	
41) Toluene-d8	10.325	98	171530	28.258	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 113.040%	
43) trans-1,3-Dichloroprop...	10.575	79	19323	24.382	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 97.520%	
47) 2-Hexanone-d5	10.922	63	12402	45.568	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 91.140%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	50776	35.439	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 141.760%#	
66) 1,2-Dichlorobenzene-d4	13.849	152	29352	29.118	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 116.480%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060324\
Data File : VW028976.D
Acq On : 03 Jun 2024 20:18
Operator : SY/MD
Sample : P2678-19
Misc : 5.48g/10mL/MSVOA_W/SOIL/A
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH688

Quant Time: Jun 04 02:22:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 04 02:14:06 2024
Response via : Initial Calibration

