

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060424\
 Data File : VW028990.D
 Acq On : 04 Jun 2024 12:58
 Operator : SY/MD
 Sample : P2678-20RE
 Misc : 5.52g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH689RE

Quant Time: Jun 05 00:48:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 05 00:46:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	303819	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	236652	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	65270	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	133226	27.603	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 110.400%	
7) Chloroethane-d5	2.893	69	105928	28.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 114.320%	
11) 1,1-Dichloroethene-d2	4.021	65	55793	26.300	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 105.200%	
21) 2-Butanone-d5	7.075	46	64578	65.485	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 130.980%	
24) Chloroform-d	7.648	84	244266	28.181	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 112.720%	
26) 1,2-Dichloroethane-d4	8.307	65	136039	29.812	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 119.240%	
32) Benzene-d6	8.276	84	475593	32.743	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 130.960%	
36) 1,2-Dichloropropane-d6	9.270	67	153470	34.848	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 139.400%#	
41) Toluene-d8	10.319	98	388068	29.939	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 119.760%	
43) trans-1,3-Dichloroprop...	10.575	79	43820	25.894	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 103.560%	
47) 2-Hexanone-d5	10.922	63	39797	68.478	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 136.960%#	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	105953	34.631	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 138.520%#	
66) 1,2-Dichlorobenzene-d4	13.849	152	77604	33.505	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 134.040%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060424\
Data File : VW028990.D
Acq On : 04 Jun 2024 12:58
Operator : SY/MD
Sample : P2678-20RE
Misc : 5.52g/10mL/MSVOA_W/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH689RE

Quant Time: Jun 05 00:48:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 05 00:46:16 2024
Response via : Initial Calibration

