

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060424\
 Data File : VW029000.D
 Acq On : 04 Jun 2024 17:17
 Operator : SY/MD
 Sample : P2710-05
 Misc : 5.35g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4B26

Quant Time: Jun 05 00:51:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 05 00:46:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	427656	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	292554	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	74964	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.368	65	180181	26.521	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.080%
7) Chloroethane-d5	2.893	69	141938	27.207	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.840%
11) 1,1-Dichloroethene-d2	4.021	65	72131	24.156	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	96.640%
21) 2-Butanone-d5	7.075	46	85552	61.633	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.260%
24) Chloroform-d	7.648	84	313500	25.695	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.800%
26) 1,2-Dichloroethane-d4	8.306	65	188411	29.333	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	117.320%
32) Benzene-d6	8.276	84	597057	33.251	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	133.000%
36) 1,2-Dichloropropane-d6	9.270	67	196649	36.120	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	144.480%#
41) Toluene-d8	10.324	98	451043	28.149	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	112.600%
43) trans-1,3-Dichloroprop...	10.574	79	51412	24.575	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.320%
47) 2-Hexanone-d5	10.922	63	50602	70.432	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	140.860%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	124982	33.045	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	132.200%#
66) 1,2-Dichlorobenzene-d4	13.848	152	67471	25.363	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.440%
Target Compounds						
16) Methylene chloride	4.923	84	22296	3.474	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060424\
Data File : VW029000.D
Acq On : 04 Jun 2024 17:17
Operator : SY/MD
Sample : P2710-05
Misc : 5.35g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4B26

Quant Time: Jun 05 00:51:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 05 00:46:16 2024
Response via : Initial Calibration

