

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060524\
 Data File : VW029008.D
 Acq On : 05 Jun 2024 09:34
 Operator : SY/MD
 Sample : VW0605SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 05 23:50:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 05 23:49:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	317197	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	298845	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	129384	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	121330	24.078	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.320%
7) Chloroethane-d5	2.899	69	93519	24.168	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.680%
11) 1,1-Dichloroethene-d2	4.027	65	54182	24.464	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	97.840%
21) 2-Butanone-d5	7.081	46	52083	50.587	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.180%
24) Chloroform-d	7.654	84	214467	23.700	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	8.465	65	81	0.017	ug/L	0.16
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.080%#
32) Benzene-d6	8.276	84	438038	23.881	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.520%
36) 1,2-Dichloropropane-d6	9.270	67	134043	24.103	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.400%
41) Toluene-d8	10.319	98	388588	23.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.960%
43) trans-1,3-Dichloroprop...	10.575	79	50105	23.447	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.800%
47) 2-Hexanone-d5	10.922	63	34765	47.370	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.740%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	96316	24.930	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	117441	25.579	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060524\
Data File : VW029008.D
Acq On : 05 Jun 2024 09:34
Operator : SY/MD
Sample : VW0605SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jun 05 23:50:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 05 23:49:22 2024
Response via : Initial Calibration

