

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060524\
 Data File : VW029012.D
 Acq On : 05 Jun 2024 11:09
 Operator : SY/MD
 Sample : P2710-10
 Misc : 5.03g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 05 23:50:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 05 23:49:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	607248	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	535086	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	217114	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	223622	23.181	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.720%
7) Chloroethane-d5	2.899	69	177439	23.953	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.800%
11) 1,1-Dichloroethene-d2	4.021	65	100170	23.625	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	94.480%
21) 2-Butanone-d5	7.075	46	96595	49.008	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.020%
24) Chloroform-d	7.648	84	406327	23.454	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	8.520	65	81	0.009	ug/L	0.21
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.040%#
32) Benzene-d6	8.276	84	844608	25.717	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.880%
36) 1,2-Dichloropropane-d6	9.276	67	257622	25.872	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.480%
41) Toluene-d8	10.325	98	759710	25.922	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	103.680%
43) trans-1,3-Dichloroprop...	10.575	79	93868	24.532	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.120%
47) 2-Hexanone-d5	10.922	63	69914	53.205	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.400%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	167912	24.273	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.080%
66) 1,2-Dichlorobenzene-d4	13.848	152	201914	26.207	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060524\
Data File : VW029012.D
Acq On : 05 Jun 2024 11:09
Operator : SY/MD
Sample : P2710-10
Misc : 5.03g/10mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jun 05 23:50:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 05 23:49:22 2024
Response via : Initial Calibration

