

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060622\
 Data File : VW023404.D
 Acq On : 06 Jun 2022 16:05
 Operator : SY/VA
 Sample : VIBLK492
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK492

Quant Time: Jun 07 01:58:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 07 01:56:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	363110	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	338177	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	155920	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	171487	30.510	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 122.040%	
7) Chloroethane-d5	2.885	69	109298	27.604	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 110.400%	
11) 1,1-Dichloroethene-d2	4.013	63	163382	19.357	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 77.440%	
21) 2-Butanone-d5	7.080	46	59247	54.656	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 109.320%	
24) Chloroform-d	7.653	84	249811	25.514	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 102.040%	
26) 1,2-Dichloroethane-d4	8.305	65	138706	26.968	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 107.880%	
32) Benzene-d6	8.275	84	500769	26.676	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 106.720%	
36) 1,2-Dichloropropane-d6	9.274	67	142244	27.034	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 108.120%	
41) Toluene-d8	10.323	98	470194	26.299	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 105.200%	
43) trans-1,3-Dichloroprop...	10.579	79	60473	25.761	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 103.040%	
47) 2-Hexanone-d5	10.920	63	44927	57.043	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 114.080%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	117033	25.996	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 104.000%	
66) 1,2-Dichlorobenzene-d4	13.853	152	153236	27.152	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 108.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060622\
Data File : VW023404.D
Acq On : 06 Jun 2022 16:05
Operator : SY/VA
Sample : VIBLK492
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK492

Quant Time: Jun 07 01:58:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 07 01:56:30 2022
Response via : Initial Calibration

