

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060622\
 Data File : VW023403.D
 Acq On : 06 Jun 2022 15:41
 Operator : SY/VA
 Sample : N3171-13
 Misc : 5.07g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jun 07 01:57:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 07 01:56:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	367543	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	338544	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	155583	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	153166	26.922	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 107.680%	
7) Chloroethane-d5	2.885	69	108856	27.161	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 108.640%	
11) 1,1-Dichloroethene-d2	4.019	63	159731	18.696	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 74.800%	
21) 2-Butanone-d5	7.074	46	61348	55.912	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 111.820%	
24) Chloroform-d	7.653	84	239342	24.150	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 96.600%	
26) 1,2-Dichloroethane-d4	8.305	65	135972	26.117	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 104.480%	
32) Benzene-d6	8.275	84	483604	25.733	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 102.920%	
36) 1,2-Dichloropropane-d6	9.274	67	137201	26.048	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 104.200%	
41) Toluene-d8	10.323	98	457496	25.561	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 102.240%	
43) trans-1,3-Dichloroprop...	10.579	79	57756	24.577	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 98.320%	
47) 2-Hexanone-d5	10.920	63	47692	60.488	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 120.980%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	118253	26.239	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 104.960%	
66) 1,2-Dichlorobenzene-d4	13.853	152	139477	24.768	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 99.080%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060622\
Data File : VW023403.D
Acq On : 06 Jun 2022 15:41
Operator : SY/VA
Sample : N3171-13
Misc : 5.07g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jun 07 01:57:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 07 01:56:30 2022
Response via : Initial Calibration

