

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060623\
 Data File : VW026110.D
 Acq On : 06 Jun 2023 16:25
 Operator : SY/MD
 Sample : 03062-06
 Misc : 5.68g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EZEL2

Quant Time: Jun 07 01:08:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 07 01:06:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	920534	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	832384	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	380007	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	259643	20.050	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.200%
7) Chloroethane-d5	2.893	69	189751	19.574	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.280%
11) 1,1-Dichloroethene-d2	4.021	65	128172	20.072	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	80.280%
21) 2-Butanone-d5	7.075	46	206325	54.725	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.460%
24) Chloroform-d	7.648	84	572724	21.246	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	8.307	65	340229	23.082	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.320%
32) Benzene-d6	8.276	84	1157131	20.659	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.640%
36) 1,2-Dichloropropane-d6	9.270	67	386227	21.766	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.080%
41) Toluene-d8	10.325	98	983439	19.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	10.575	79	146342	20.812	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.240%
47) 2-Hexanone-d5	10.922	63	139382	49.816	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.640%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	323442	25.049	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	13.848	152	323822	22.263	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060623\
Data File : VW026110.D
Acq On : 06 Jun 2023 16:25
Operator : SY/MD
Sample : 03062-06
Misc : 5.68g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EZEL2

Quant Time: Jun 07 01:08:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 07 01:06:46 2023
Response via : Initial Calibration

