

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060623\
 Data File : VW026111.D
 Acq On : 06 Jun 2023 16:49
 Operator : SY/MD
 Sample : 03062-07
 Misc : 6.19g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EZEL3

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/07/2023
 Supervised By :Mahesh Dadoda 06/07/2023

Quant Time: Jun 07 01:08:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 07 01:06:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	4051	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.636	117	3044	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.568	152	992m	25.000	ug/L	0.01
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	1650m	28.953	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 115.800%	
7) Chloroethane-d5	2.905	69	1691	39.638	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	= 158.560%#	
11) 1,1-Dichloroethene-d2	4.033	65	1246m	44.340	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 177.360%#	
21) 2-Butanone-d5	0.000	46	0d	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	= 0.000%#	
24) Chloroform-d	7.648	84	3714m	31.308	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 125.240%	
26) 1,2-Dichloroethane-d4	8.313	65	2254m	34.748	ug/L	0.01
Spiked Amount	25.000	Range	70 - 130	Recovery	= 139.000%#	
32) Benzene-d6	8.270	84	5955	29.073	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 116.280%	
36) 1,2-Dichloropropane-d6	9.276	67	2176	33.533	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 134.120%#	
41) Toluene-d8	10.319	98	3397	18.701	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 74.800%	
43) trans-1,3-Dichloroprop...	10.569	79	492m	19.133	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 76.520%	
47) 2-Hexanone-d5	10.971	63	105	10.262	ug/L	0.05
Spiked Amount	50.000	Range	20 - 135	Recovery	= 20.520%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	1196	25.328	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 101.320%	
66) 1,2-Dichlorobenzene-d4	13.855	152	662m	17.435	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 69.760%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060623\
Data File : VW026111.D
Acq On : 06 Jun 2023 16:49
Operator : SY/MD
Sample : 03062-07
Misc : 6.19g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EZEL3

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 06/07/2023
Supervised By : Mahesh Dadoda 06/07/2023

Quant Time: Jun 07 01:08:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 07 01:06:46 2023
Response via : Initial Calibration

