

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060624\
 Data File : VW029038.D
 Acq On : 06 Jun 2024 15:20
 Operator : SY/MD
 Sample : P2696-06
 Misc : 5.30g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BN4

Quant Time: Jun 07 00:25:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 07 00:21:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	569274	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	413423	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	106181	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	176535	19.520	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.080%
7) Chloroethane-d5	2.893	69	143312	20.636	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.560%
11) 1,1-Dichloroethene-d2	4.021	65	76821	19.326	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.320%
21) 2-Butanone-d5	7.075	46	91340	49.433	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.860%
24) Chloroform-d	7.648	84	323852	19.941	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.760%
26) 1,2-Dichloroethane-d4	8.307	65	189246	22.133	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.520%
32) Benzene-d6	8.270	84	643764	25.370	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.480%
36) 1,2-Dichloropropane-d6	9.270	67	203724	26.480	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.920%
41) Toluene-d8	10.319	98	494185	21.824	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.280%
43) trans-1,3-Dichloroprop...	10.575	79	65146	22.036	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.160%
47) 2-Hexanone-d5	10.922	63	63558	62.602	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.200%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	113159	21.172	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.680%
66) 1,2-Dichlorobenzene-d4	13.849	152	65698	17.436	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	69.760%#
Target Compounds						Qvalue
13) Acetone	4.119	43	96641	49.908	ug/L	100
22) 2-Butanone	7.173	43	50902	21.815	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060624\
Data File : VW029038.D
Acq On : 06 Jun 2024 15:20
Operator : SY/MD
Sample : P2696-06
Misc : 5.30g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BN4

Quant Time: Jun 07 00:25:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 07 00:21:46 2024
Response via : Initial Calibration

