

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060724\
 Data File : VW029064.D
 Acq On : 07 Jun 2024 15:53
 Operator : SY/MD
 Sample : P2696-26RE
 Misc : 4.57g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BP0RE

Quant Time: Jun 07 23:10:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 07 23:05:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	532185	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	336876	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	73412	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	160958	19.038	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.160%
7) Chloroethane-d5	2.899	69	136096	20.963	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.840%
11) 1,1-Dichloroethene-d2	4.021	65	71325	19.194	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.760%
21) 2-Butanone-d5	7.081	46	53015	30.691	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.380%
24) Chloroform-d	7.648	84	303256	19.974	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.880%
26) 1,2-Dichloroethane-d4	8.307	65	163728	20.483	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.920%
32) Benzene-d6	8.276	84	606070	29.312	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	117.240%
36) 1,2-Dichloropropane-d6	9.276	67	185503	29.590	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	118.360%
41) Toluene-d8	10.319	98	456142	24.721	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.880%
43) trans-1,3-Dichloroprop...	10.575	79	49639	20.606	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.440%
47) 2-Hexanone-d5	10.922	63	38455	46.483	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.960%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	86464	19.853	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.400%
66) 1,2-Dichlorobenzene-d4	13.848	152	51229	19.665	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060724\
Data File : VW029064.D
Acq On : 07 Jun 2024 15:53
Operator : SY/MD
Sample : P2696-26RE
Misc : 4.57g/10mL/MSVOA_W/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BP0RE

Quant Time: Jun 07 23:10:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 07 23:05:54 2024
Response via : Initial Calibration

