

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060823\
 Data File : VW026151.D
 Acq On : 08 Jun 2023 23:06
 Operator : SY/MD
 Sample : VIBLK467
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK467

Quant Time: Jun 09 01:40:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 09 01:33:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	820002	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	738398	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	351328	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	213022	18.466	ug/L	-0.01
Spiked Amount 25.000	Range 30 - 150		Recovery =	73.880%		
7) Chloroethane-d5	2.899	69	189893	21.990	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	87.960%		
11) 1,1-Dichloroethene-d2	4.021	65	109148	19.189	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	76.760%		
21) 2-Butanone-d5	7.075	46	155147	46.196	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	92.400%		
24) Chloroform-d	7.648	84	484742	20.187	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.760%		
26) 1,2-Dichloroethane-d4	8.307	65	283579	21.597	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	86.400%		
32) Benzene-d6	8.276	84	966459	19.451	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	77.800%		
36) 1,2-Dichloropropane-d6	9.270	67	312791	19.871	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	79.480%		
41) Toluene-d8	10.325	98	822269	18.661	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	74.640%		
43) trans-1,3-Dichloroprop...	10.575	79	112169	17.983	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	71.920%		
47) 2-Hexanone-d5	10.922	63	98116	39.531	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	79.060%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	234138	20.441	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	81.760%		
66) 1,2-Dichlorobenzene-d4	13.848	152	255751	19.019	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	76.080%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060823\
Data File : VW026151.D
Acq On : 08 Jun 2023 23:06
Operator : SY/MD
Sample : VIBLK467
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK467

Quant Time: Jun 09 01:40:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 09 01:33:49 2023
Response via : Initial Calibration

