

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060923\
 Data File : VW026162.D
 Acq On : 09 Jun 2023 11:45
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK470

Quant Time: Jun 09 22:54:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 09 22:52:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	841179	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	753329	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	349765	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	233898	19.766	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.080%
7) Chloroethane-d5	2.899	69	200019	22.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.320%
11) 1,1-Dichloroethene-d2	4.021	65	119044	20.402	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.600%
21) 2-Butanone-d5	7.075	46	154288	44.784	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.560%
24) Chloroform-d	7.648	84	493425	20.031	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.120%
26) 1,2-Dichloroethane-d4	8.307	65	287951	21.378	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.520%
32) Benzene-d6	8.270	84	1003577	19.798	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.200%
36) 1,2-Dichloropropane-d6	9.270	67	311672	19.408	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.640%
41) Toluene-d8	10.325	98	839787	18.681	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	74.720%
43) trans-1,3-Dichloroprop...	10.575	79	114959	18.065	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.240%
47) 2-Hexanone-d5	10.922	63	108825	42.976	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.960%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	233731	20.001	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.000%
66) 1,2-Dichlorobenzene-d4	13.848	152	261056	19.500	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060923\
Data File : VW026162.D
Acq On : 09 Jun 2023 11:45
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK470

Quant Time: Jun 09 22:54:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 09 22:52:55 2023
Response via : Initial Calibration

