

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061024\
 Data File : VW029074.D
 Acq On : 10 Jun 2024 10:23
 Operator : SY/MD
 Sample : P2710-20
 Misc : 5.03g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jun 11 01:19:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 11 01:19:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	312211	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	282182	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	134877	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	101939	20.553	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.200%
7) Chloroethane-d5	2.899	69	83062	21.808	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.240%
11) 1,1-Dichloroethene-d2	4.027	65	46263	21.222	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	84.880%
21) 2-Butanone-d5	7.081	46	58341	57.571	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.140%
24) Chloroform-d	7.648	84	203591	22.857	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.440%
26) 1,2-Dichloroethane-d4	8.307	65	116868	24.922	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.680%
32) Benzene-d6	8.276	84	412439	23.813	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.240%
36) 1,2-Dichloropropane-d6	9.276	67	129901	24.737	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.960%
41) Toluene-d8	10.324	98	355721	23.016	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.080%
43) trans-1,3-Dichloroprop...	10.574	79	49375	24.469	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.880%
47) 2-Hexanone-d5	10.922	63	39648	57.214	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.420%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	100897	27.658	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	110.640%
66) 1,2-Dichlorobenzene-d4	13.848	152	123106	25.721	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061024\
Data File : VW029074.D
Acq On : 10 Jun 2024 10:23
Operator : SY/MD
Sample : P2710-20
Misc : 5.03g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jun 11 01:19:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 11 01:19:14 2024
Response via : Initial Calibration

