

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061024\
 Data File : VW029088.D
 Acq On : 10 Jun 2024 16:50
 Operator : SY/MD
 Sample : P2776-16DL 2X
 Misc : 5.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4DG7DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024
 Supervised By :Mahesh Dadoda 06/11/2024

Quant Time: Jun 11 01:23:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 11 01:19:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	314716	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	286971	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	141737	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	96524	19.306	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	77.240%		
7) Chloroethane-d5	2.899	69	75580	19.686	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	78.760%		
11) 1,1-Dichloroethene-d2	4.027	65	44890	20.428	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	81.720%		
21) 2-Butanone-d5	7.075	46	60139	58.873	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	117.740%		
24) Chloroform-d	7.648	84	193478	21.549	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	86.200%		
26) 1,2-Dichloroethane-d4	8.307	65	109397	23.143	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	92.560%		
32) Benzene-d6	8.276	84	391441	22.224	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	88.880%		
36) 1,2-Dichloropropane-d6	9.270	67	122409	22.921	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	91.680%		
41) Toluene-d8	10.325	98	349923	22.263	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	89.040%		
43) trans-1,3-Dichloroprop...	10.575	79	47458	23.127	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	92.520%		
47) 2-Hexanone-d5	10.922	63	43474	61.688	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	123.380%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	98941	26.669	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	106.680%		
66) 1,2-Dichlorobenzene-d4	13.848	152	118300	23.520	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	94.080%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	4.051	96	202200m	50.829	ug/L	
13) Acetone	4.118	43	108364	101.227	ug/L	100
16) Methylene chloride	4.923	84	114032	24.142	ug/L	98
17) trans-1,2-Dichloroethene	5.429	96	56893	13.117	ug/L	98
18) Methyl tert-butyl Ether	5.423	73	67282	11.124	ug/L	98
20) cis-1,2-Dichloroethene	7.167	96	74162	16.310	ug/L	98
22) 2-Butanone	7.167	43	89365	69.278	ug/L	99
27) 1,2-Dichloroethane	8.398	62	105021	18.922	ug/L	98
31) Carbon tetrachloride	8.069	117	229426	37.882	ug/L	96
33) Benzene	8.325	78	496161	27.806	ug/L	100
34) Trichloroethene	9.093	95	233202	49.711	ug/L	92
37) 1,2-Dichloropropane	9.367	63	120448	25.777	ug/L	100
38) Bromodichloromethane	9.648	83	166076	29.142	ug/L	97
39) cis-1,3-Dichloropropene	10.069	75	197804	29.128	ug/L	98
40) 4-Methyl-2-pentanone	10.209	43	299258	125.911	ug/L	99
44) trans-1,3-Dichloropropene	10.605	75	107582	19.651	ug/L	96
45) 1,1,2-Trichloroethane	10.782	97	75633	23.966	ug/L	94
46) Tetrachloroethene	10.861	164	94595	26.541	ug/L	87
48) 2-Hexanone	10.965	43	121016	67.911	ug/L	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061024\
 Data File : VW029088.D
 Acq On : 10 Jun 2024 16:50
 Operator : SY/MD
 Sample : P2776-16DL 2X
 Misc : 5.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4DG7DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024
 Supervised By :Mahesh Dadoda 06/11/2024

Quant Time: Jun 11 01:23:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 11 01:19:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

49) Dibromochloromethane	11.129	129	54456	16.337	ug/L	95
50) 1,2-Dibromoethane	11.233	107	28318	9.830	ug/L	97
52) Ethylbenzene	11.727	91	923467	44.588	ug/L	98
53) m,p-Xylene	11.837	106	299538	39.047	ug/L	92
54) o-Xylene	12.166	106	35266	4.965	ug/L	96
57) 1,1,2,2-Tetrachloroethane	12.714	83	39513	11.091	ug/L	97
59) Bromoform	12.349	173	47794	26.223	ug/L	97
61) 1,2,3-Trichloropropane	12.763	75	104081	38.930	ug/L	98
64) 1,3-Dichlorobenzene	13.495	146	313641	36.385	ug/L	91
65) 1,4-Dichlorobenzene	13.574	146	125961	13.991	ug/L	95
67) 1,2-Dichlorobenzene	13.867	146	347858	45.024	ug/L	97
68) 1,2-Dibromo-3-chloropr...	14.476	75	8267	14.507	ug/L #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061024\
Data File : VW029088.D
Acq On : 10 Jun 2024 16:50
Operator : SY/MD
Sample : P2776-16DL 2X
Misc : 5.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024

Quant Time: Jun 11 01:23:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 11 01:19:14 2024
Response via : Initial Calibration









#12
1,1-Dichloroethene
Concen: 50.829 ug/L m
RT: 4.051 min Scan# 355
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Tgt Ion: 96 Resp: 202200
Ion Ratio Lower Upper
96 100
61 168.3 132.2 245.4
63 84.8 108.4 201.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#13
Acetone
Concen: 101.227 ug/L
RT: 4.118 min Scan# 366
Delta R.T. -0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50


Tgt Ion: 43 Resp: 108364
Ion Ratio Lower Upper
43 100
58 29.7 0.0 59.4





#16
Methylene chloride
Concen: 24.142 ug/L
RT: 4.923 min Scan# 498
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50




Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL



Tgt Ion: 84 Resp: 11403
Ion Ratio Lower Upper
84 100
86 62.5 43.8 81.4
49 134.3 91.3 169.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#17
trans-1,2-Dichloroethene
Concen: 13.117 ug/L
RT: 5.429 min Scan# 581
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Tgt Ion: 96 Resp: 56893
Ion Ratio Lower Upper
96 100
61 150.0 104.6 194.2
98 65.9 44.0 81.8









#18
Methyl tert-butyl Ether
Concen: 11.124 ug/L
RT: 5.423 min Scan# 581
Delta R.T. -0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Tgt Ion: 73 Resp: 6728
Ion Ratio Lower Upper
73 100
43 23.5 17.8 26.6
57 23.7 18.4 27.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#20
cis-1,2-Dichloroethene
Concen: 16.310 ug/L
RT: 7.167 min Scan# 866
Delta R.T. -0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50


Tgt Ion: 96 Resp: 74162
Ion Ratio Lower Upper
96 100
61 133.5 92.9 172.5
98 60.3 45.2 84.0





#22
2-Butanone
Concen: 69.278 ug/L
RT: 7.167 min Scan# 80
Delta R.T. -0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50




Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL



Tgt Ion: 43 Resp: 8936
Ion Ratio Lower Upper
43 100
72 24.4 12.5 37.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#27
1,2-Dichloroethane
Concen: 18.922 ug/L
RT: 8.398 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50


Tgt Ion: 62 Resp: 105021
Ion Ratio Lower Upper
62 100
98 9.6 7.1 10.7





#31
Carbon tetrachloride
Concen: 37.882 ug/L
RT: 8.069 min Scan# 1015
Delta R.T. -0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50




Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL



Tgt Ion:117 Resp: 229420
Ion Ratio Lower Upper
117 100
119 94.5 78.8 118.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#33
Benzene
Concen: 27.806 ug/L
RT: 8.325 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50


Tgt Ion: 78 Resp: 496161





#34
Trichloroethene
Concen: 49.711 ug/L
RT: 9.093 min Scan# 1182
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50




Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL




Tgt Ion: 95 Resp: 23320
Ion Ratio Lower Upper
95 100
97 62.5 41.9 77.7
132 96.9 60.1 111.5
130 97.9 64.2 119.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024




#37
1,2-Dichloropropane
Concen: 25.777 ug/L
RT: 9.367 min Scan# 1227
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50



Tgt Ion: 63 Resp: 120448
Ion Ratio Lower Upper
63 100
112 4.4 3.6 5.4





#38

Bromodichloromethane

Concen: 29.142 ug/L

RT: 9.648 min Scan# 1273

Delta R.T. 0.006 min

Lab File: VW029088.D

Acq: 10 Jun 2024 16:50

Instrument :

MSVOA_W

ClientSampleId :





A4DG7DL

Tgt Ion: 83 Resp: 166070
 Ion Ratio Lower Upper
 83 100
 85 60.4 43.9 81.5
 127 8.0 5.4 8.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024
 Supervised By :Mahesh Dadoda 06/11/2024



#39

cis-1,3-Dichloropropene

Concen: 29.128 ug/L




RT: 10.069 min Scan# 1342


Delta R.T. -0.000 min

Lab File: VW029088.D

Acq: 10 Jun 2024 16:50


Tgt Ion: 75 Resp: 197804
 Ion Ratio Lower Upper
 75 100
 77 31.0 21.0 39.0





#40
4-Methyl-2-pentanone
Concen: 125.911 ug/L
RT: 10.209 min Scan# 1365
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50




Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL



Tgt Ion: 43 Resp: 299258
Ion Ratio Lower Upper
43 100
58 37.1 30.3 45.5
100 13.7 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#44
trans-1,3-Dichloropropene
Concen: 19.651 ug/L
RT: 10.605 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Tgt Ion: 75 Resp: 107582
Ion Ratio Lower Upper
75 100
77 29.3 22.0 40.8





#45

1,1,2-Trichloroethane

Concen: 23.966 ug/L

RT: 10.782 min Scan# 1459

Delta R.T. -0.000 min

Lab File: VW029088.D

Acq: 10 Jun 2024 16:50

Instrument :

MSVOA_W

ClientSampleId :

A4DG7DL





Tgt Ion:	97	Resp:	7563
Ion	Ratio	Lower	Upper
97	100		
99	61.9	41.8	77.6
83	80.9	62.4	115.8
85	55.8	41.5	77.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024

Supervised By :Mahesh Dadoda 06/11/2024



#46

Tetrachloroethene

Concen: 26.541 ug/L




RT: 10.861 min Scan# 1472


Delta R.T. -0.000 min

Lab File: VW029088.D

Acq: 10 Jun 2024 16:50

Tgt Ion:	164	Resp:	94595
Ion	Ratio	Lower	Upper
164	100		
129	97.8	69.4	128.8
131	86.4	71.0	131.9
166	111.9	96.0	178.4





#48

2-Hexanone

Concen: 67.911 ug/L

RT: 10.965 min Scan# 1489

Delta R.T. -0.000 min

Lab File: VW029088.D


Acq: 10 Jun 2024 16:50

Instrument :

MSVOA_W

ClientSampleId :

A4DG7DL



Tgt Ion: 43 Resp: 121010

Ion Ratio Lower Upper

43 100

58 52.2 41.4 62.0

57 18.3 14.2 21.4




100 11.3 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024

Supervised By :Mahesh Dadoda 06/11/2024



#49

Dibromochloromethane

Concen: 16.337 ug/L

RT: 11.129 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VW029088.D




Acq: 10 Jun 2024 16:50


Tgt Ion:129 Resp: 54456

Ion Ratio Lower Upper

129 100

127 84.7 56.5 104.9





#50

1,2-Dibromoethane

Concen: 9.830 ug/L

RT: 11.233 min Scan# 1533

Delta R.T. -0.000 min

Lab File: VW029088.D

Acq: 10 Jun 2024 16:50

Instrument :

MSVOA_W

ClientSampleId :





A4DG7DL

Tgt Ion:	107	Resp:	28318
Ion Ratio	Lower	Upper	
107	100		
109	99.9	67.9	126.1
188	3.7	3.0	4.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#52

Ethylbenzene

Concen: 44.588 ug/L




RT: 11.727 min Scan# 1614


Delta R.T. -0.000 min

Lab File: VW029088.D

Acq: 10 Jun 2024 16:50

Tgt Ion:	91	Resp:	923467
Ion Ratio	Lower	Upper	
91	100		
106	29.7	19.9	37.0









#53
m,p-Xylene
Concen: 39.047 ug/L
RT: 11.837 min Scan# 1632
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Tgt Ion:106 Resp: 299538
Ion Ratio Lower Upper
106 100
91 203.6 152.0 282.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#54
o-Xylene
Concen: 4.965 ug/L
RT: 12.166 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Tgt Ion:106 Resp: 35266
Ion Ratio Lower Upper
106 100
91 224.0 152.5 283.3









#57
1,1,2,2-Tetrachloroethane
Concen: 11.091 ug/L
RT: 12.714 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Tgt Ion: 83 Resp: 3951
Ion Ratio Lower Upper
83 100
85 60.0 43.8 81.3
131 10.1 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#59
Bromoform
Concen: 26.223 ug/L
RT: 12.349 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Tgt Ion:173 Resp: 47794
Ion Ratio Lower Upper
173 100
175 50.2 38.1 57.1
254 10.2 8.2 12.2









#61
1,2,3-Trichloropropane
Concen: 38.930 ug/L
RT: 12.763 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Tgt Ion: 75 Resp: 10408
Ion Ratio Lower Upper
75 100
77 32.1 26.3 39.5
110 37.9 29.4 44.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#64
1,3-Dichlorobenzene
Concen: 36.385 ug/L
RT: 13.495 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Tgt Ion:146 Resp: 313641
Ion Ratio Lower Upper
146 100
111 46.6 29.4 54.6
148 68.4 42.5 78.9









#65
1,4-Dichlorobenzene
Concen: 13.991 ug/L
RT: 13.574 min Scan# 1917
Delta R.T. -0.000 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Instrument :
MSVOA_W
ClientSampleId :
A4DG7DL

Tgt Ion:146 Resp: 12596
Ion Ratio Lower Upper
146 100
111 41.5 27.5 51.1
148 61.1 39.5 73.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/11/2024
Supervised By :Mahesh Dadoda 06/11/2024



#67
1,2-Dichlorobenzene
Concen: 45.024 ug/L
RT: 13.867 min Scan# 1965
Delta R.T. 0.006 min
Lab File: VW029088.D
Acq: 10 Jun 2024 16:50

Tgt Ion:146 Resp: 347858
Ion Ratio Lower Upper
146 100
111 42.4 32.8 49.2
148 62.8 48.2 72.4





#68
 1,2-Dibromo-3-chloropropane
 Concen: 14.507 ug/L
 RT: 14.476 min Scan# 2065
 Delta R.T. -0.000 min
 Lab File: VW029088.D
 Acq: 10 Jun 2024 16:50

Instrument :
 MSVOA_W
 ClientSampleId :
 A4DG7DL

Tgt Ion: 75 Resp: 826
 Ion Ratio Lower Upper
 75 100
 155 81.8 52.3 78.5
 157 94.5 71.0 106.4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/11/2024
 Supervised By :Mahesh Dadoda 06/11/2024

