

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061223\
 Data File : VW026175.D
 Acq On : 12 Jun 2023 11:02
 Operator : SY/MD
 Sample : 03078-06
 Misc : 5.05g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jun 13 01:21:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 13 01:20:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	782525	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	685564	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	316130	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	233200	21.184	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.720%
7) Chloroethane-d5	2.899	69	203044	24.639	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.560%
11) 1,1-Dichloroethene-d2	4.021	65	119655	22.043	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.160%
21) 2-Butanone-d5	7.075	46	171648	53.557	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.120%
24) Chloroform-d	7.648	84	503609	21.977	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.920%
26) 1,2-Dichloroethane-d4	8.307	65	301571	24.067	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.280%
32) Benzene-d6	8.276	84	1020452	22.121	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.480%
36) 1,2-Dichloropropane-d6	9.270	67	331049	22.652	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.600%
41) Toluene-d8	10.325	98	890405	21.765	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.040%
43) trans-1,3-Dichloroprop...	10.575	79	124452	21.489	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.960%
47) 2-Hexanone-d5	10.922	63	116343	50.487	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	253140	23.803	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	13.854	152	272652	22.533	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061223\
Data File : VW026175.D
Acq On : 12 Jun 2023 11:02
Operator : SY/MD
Sample : 03078-06
Misc : 5.05g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jun 13 01:21:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 13 01:20:44 2023
Response via : Initial Calibration

