

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061224\
 Data File : VW029102.D
 Acq On : 12 Jun 2024 13:15
 Operator : SY/MD
 Sample : P2776-21
 Misc : 4.71g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4C74

Quant Time: Jun 13 00:49:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 13 00:47:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	531732	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	469326	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	204031	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	161394	19.106	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.440%
7) Chloroethane-d5	2.899	69	135285	20.856	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.440%
11) 1,1-Dichloroethene-d2	4.021	65	74818	20.151	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	80.600%
21) 2-Butanone-d5	7.075	46	81010	46.938	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.880%
24) Chloroform-d	7.654	84	337612	22.255	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.040%
26) 1,2-Dichloroethane-d4	8.307	65	179984	22.536	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.160%
32) Benzene-d6	8.276	84	669858	23.254	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	9.270	67	206723	23.669	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.680%
41) Toluene-d8	10.325	98	579760	22.554	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.200%
43) trans-1,3-Dichloroprop...	10.574	79	75575	22.519	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.080%
47) 2-Hexanone-d5	10.922	63	58780	50.999	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	143587	23.665	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.640%
66) 1,2-Dichlorobenzene-d4	13.854	152	174554	24.109	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.440%
Target Compounds						
16) Methylene chloride	4.923	84	19780	2.479	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061224\
Data File : VW029102.D
Acq On : 12 Jun 2024 13:15
Operator : SY/MD
Sample : P2776-21
Misc : 4.71g/10mL/MSVOA_W/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4C74

Quant Time: Jun 13 00:49:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 13 00:47:39 2024
Response via : Initial Calibration

