

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061224\
 Data File : VW029112.D
 Acq On : 12 Jun 2024 17:14
 Operator : SY/MD
 Sample : P2844-05
 Misc : 5.02g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4C83

Quant Time: Jun 13 00:51:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jun 13 00:47:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	484363	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	452509	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	214896	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	146741	19.070	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.280%
7) Chloroethane-d5	2.899	69	125253	21.198	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.800%
11) 1,1-Dichloroethene-d2	4.021	65	69792	20.636	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.560%
21) 2-Butanone-d5	7.075	46	84472	53.730	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.460%
24) Chloroform-d	7.648	84	324503	23.483	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.920%
26) 1,2-Dichloroethane-d4	8.307	65	178371	24.518	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.080%
32) Benzene-d6	8.276	84	642313	23.127	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.520%
36) 1,2-Dichloropropane-d6	9.276	67	205589	24.414	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.640%
41) Toluene-d8	10.325	98	568692	22.945	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.800%
43) trans-1,3-Dichloroprop...	10.575	79	73792	22.805	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.200%
47) 2-Hexanone-d5	10.922	63	61733	55.552	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.100%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	149828	25.611	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.440%
66) 1,2-Dichlorobenzene-d4	13.849	152	186573	24.466	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.880%
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	15365	2.114	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061224\
Data File : VW029112.D
Acq On : 12 Jun 2024 17:14
Operator : SY/MD
Sample : P2844-05
Misc : 5.02g/10mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4C83

Quant Time: Jun 13 00:51:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jun 13 00:47:39 2024
Response via : Initial Calibration

