

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061322\
 Data File : VW023473.D
 Acq On : 13 Jun 2022 12:24
 Operator : SY/VA
 Sample : N3226-13
 Misc : 6.30g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4N3

Quant Time: Jun 14 01:34:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 14 01:31:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	450605	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	379265	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	117334	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	146562	20.473	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.880%
7) Chloroethane-d5	2.879	69	106933	20.320	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.280%
11) 1,1-Dichloroethene-d2	4.013	63	154916	15.504	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.000%
21) 2-Butanone-d5	7.074	46	58646	40.124	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.240%
24) Chloroform-d	7.647	84	273240	22.028	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.120%
26) 1,2-Dichloroethane-d4	8.305	65	152908	22.889	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.560%
32) Benzene-d6	8.275	84	500668	24.897	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	9.275	67	155419	26.233	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.920%
41) Toluene-d8	10.323	98	443944	22.602	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.400%
43) trans-1,3-Dichloroprop...	10.579	79	55402	20.488	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.960%
47) 2-Hexanone-d5	10.927	63	40605	42.891	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.780%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	125328	22.649	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	97244	22.958	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.840%
Target Compounds						
29) Cyclohexane	7.952	56	6414	0.840	ug/L	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061322\
Data File : VW023473.D
Acq On : 13 Jun 2022 12:24
Operator : SY/VA
Sample : N3226-13
Misc : 6.30g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4N3

Quant Time: Jun 14 01:34:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 14 01:31:49 2022
Response via : Initial Calibration

