

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
 Data File : VW029125.D
 Acq On : 13 Jun 2024 15:17
 Operator : SY/MD
 Sample : P2844-11
 Misc : 5.01g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4C63

Quant Time: Jun 14 05:22:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 14 05:21:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	472469	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	423681	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	179050	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	153079	20.395	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.560%
7) Chloroethane-d5	2.899	69	131895	22.884	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.520%
11) 1,1-Dichloroethene-d2	4.021	65	70096	21.248	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.000%
21) 2-Butanone-d5	7.075	46	97459	63.551	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	127.100%
24) Chloroform-d	7.648	84	324577	24.080	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.320%
26) 1,2-Dichloroethane-d4	8.307	65	192228	27.088	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.360%
32) Benzene-d6	8.270	84	631007	24.265	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.080%
36) 1,2-Dichloropropane-d6	9.276	67	203211	25.774	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.080%
41) Toluene-d8	10.318	98	552038	23.789	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.160%
43) trans-1,3-Dichloroprop...	10.575	79	75632	24.964	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.840%
47) 2-Hexanone-d5	10.922	63	69624	66.916	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	133.840%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	153312	27.990	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.960%
66) 1,2-Dichlorobenzene-d4	13.848	152	153999	24.238	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
Data File : VW029125.D
Acq On : 13 Jun 2024 15:17
Operator : SY/MD
Sample : P2844-11
Misc : 5.01g/10mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4C63

Quant Time: Jun 14 05:22:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 14 05:21:04 2024
Response via : Initial Calibration

