

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
 Data File : VW029134.D
 Acq On : 13 Jun 2024 18:54
 Operator : SY/MD
 Sample : P2870-07
 Misc : 4.80g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AA8

Quant Time: Jun 14 05:24:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 14 05:21:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	396733	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	327379	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	116592	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	129838	20.601	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.400%
7) Chloroethane-d5	2.899	69	112620	23.270	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.080%
11) 1,1-Dichloroethene-d2	4.021	65	57185	20.643	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.560%
21) 2-Butanone-d5	7.075	46	70916	55.071	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.140%
24) Chloroform-d	7.648	84	279314	24.678	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.720%
26) 1,2-Dichloroethane-d4	8.307	65	168062	28.204	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.800%
32) Benzene-d6	8.270	84	528629	26.308	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.240%
36) 1,2-Dichloropropane-d6	9.270	67	173774	28.523	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	114.080%
41) Toluene-d8	10.319	98	425195	23.713	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.840%
43) trans-1,3-Dichloroprop...	10.575	79	60649	25.907	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.640%
47) 2-Hexanone-d5	10.922	63	45354	56.413	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.820%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	118732	28.053	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.200%
66) 1,2-Dichlorobenzene-d4	13.849	152	105008	25.380	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.520%
Target Compounds						
					Qvalue	
13) Acetone	4.125	43	27361	20.275	ug/L	97
16) Methylene chloride	4.917	84	44620	7.494	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
Data File : VW029134.D
Acq On : 13 Jun 2024 18:54
Operator : SY/MD
Sample : P2870-07
Misc : 4.80g/10mL/MSVOA_W/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AA8

Quant Time: Jun 14 05:24:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 14 05:21:04 2024
Response via : Initial Calibration

