

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
 Data File : VW029137.D
 Acq On : 13 Jun 2024 20:06
 Operator : SY/MD
 Sample : P2870-10
 Misc : 3.59g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB1

Quant Time: Jun 14 05:25:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 14 05:21:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	419349	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	313933	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	82005	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	142351	21.368	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.480%
7) Chloroethane-d5	2.899	69	128969	25.210	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.840%
11) 1,1-Dichloroethene-d2	4.021	65	65426	22.344	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	89.360%
21) 2-Butanone-d5	7.075	46	57741	42.421	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.840%
24) Chloroform-d	7.648	84	308382	25.777	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.120%
26) 1,2-Dichloroethane-d4	8.301	65	171588	27.243	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.960%
32) Benzene-d6	8.270	84	568265	29.492	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	117.960%
36) 1,2-Dichloropropane-d6	9.270	67	186332	31.895	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	127.560%#
41) Toluene-d8	10.325	98	464397	27.008	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	108.040%
43) trans-1,3-Dichloroprop...	10.575	79	49647	22.116	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.480%
47) 2-Hexanone-d5	10.922	63	29687	38.507	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.020%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	108473	26.727	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.920%
66) 1,2-Dichlorobenzene-d4	13.848	152	74242	25.513	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.040%
Target Compounds						
					Qvalue	
13) Acetone	4.125	43	8561	6.002	ug/L	100
16) Methylene chloride	4.917	84	45453	7.222	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
Data File : VW029137.D
Acq On : 13 Jun 2024 20:06
Operator : SY/MD
Sample : P2870-10
Misc : 3.59g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AB1

Quant Time: Jun 14 05:25:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 14 05:21:04 2024
Response via : Initial Calibration

