

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
 Data File : VW029139.D
 Acq On : 13 Jun 2024 20:54
 Operator : SY/MD
 Sample : P2870-12
 Misc : 2.41g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB3

Quant Time: Jun 14 05:25:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 14 05:21:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	411860	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	321693	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	107911	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	136888	20.921	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.680%
7) Chloroethane-d5	2.899	69	122353	24.352	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.400%
11) 1,1-Dichloroethene-d2	4.021	65	58554	20.361	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.440%
21) 2-Butanone-d5	7.075	46	52308	39.129	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.260%
24) Chloroform-d	7.648	84	290478	24.722	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.880%
26) 1,2-Dichloroethane-d4	8.307	65	159934	25.854	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	8.276	84	529942	26.840	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.360%
36) 1,2-Dichloropropane-d6	9.270	67	175044	29.240	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	116.960%
41) Toluene-d8	10.319	98	438116	24.865	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.480%
43) trans-1,3-Dichloroprop...	10.575	79	37055	16.108	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.440%
47) 2-Hexanone-d5	10.922	63	24319	30.783	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.560%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	104573	25.144	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.560%
66) 1,2-Dichlorobenzene-d4	13.849	152	81840	21.372	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.480%
Target Compounds						
					Qvalue	
13) Acetone	4.125	43	5310	3.790	ug/L	88
16) Methylene chloride	4.923	84	29728	4.809	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061324\
Data File : VW029139.D
Acq On : 13 Jun 2024 20:54
Operator : SY/MD
Sample : P2870-12
Misc : 2.41g/10mL/MSVOA_W/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AB3

Quant Time: Jun 14 05:25:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 14 05:21:04 2024
Response via : Initial Calibration

