

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061423\
 Data File : VW026186.D
 Acq On : 14 Jun 2023 18:27
 Operator : SY/MD
 Sample : 03234-10
 Misc : 4.40g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YBTX6

Quant Time: Jun 14 23:27:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jun 14 23:24:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	849823	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	692051	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.562	152	260498	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	231466	19.361	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.440%
7) Chloroethane-d5	2.899	69	210746	23.548	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	4.021	65	119759	20.315	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.280%
21) 2-Butanone-d5	7.075	46	195903	56.284	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.560%
24) Chloroform-d	7.648	84	543841	21.853	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.400%
26) 1,2-Dichloroethane-d4	8.508	65	112	0.008	ug/L	0.21
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.040%#
32) Benzene-d6	8.276	84	1086942	23.341	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.360%
36) 1,2-Dichloropropane-d6	9.276	67	360664	24.447	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.800%
41) Toluene-d8	10.324	98	899234	21.775	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.080%
43) trans-1,3-Dichloroprop...	10.574	79	116701	19.962	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.840%
47) 2-Hexanone-d5	10.922	63	126809	54.513	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.020%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	277236	25.824	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.280%
66) 1,2-Dichlorobenzene-d4	13.854	152	244160	24.488	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061423\
Data File : VW026186.D
Acq On : 14 Jun 2023 18:27
Operator : SY/MD
Sample : 03234-10
Misc : 4.40g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YBTX6

Quant Time: Jun 14 23:27:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM053023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jun 14 23:24:58 2023
Response via : Initial Calibration

