

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061424\
 Data File : VW029149.D
 Acq On : 14 Jun 2024 10:39
 Operator : SY/MD
 Sample : P2870-18
 Misc : 5.84g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB9

Quant Time: Jun 15 00:21:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jun 15 00:20:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	450625	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	347637	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	99149	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	145423	20.314	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	81.240%		
7) Chloroethane-d5	2.899	69	128539	23.382	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	93.520%		
11) 1,1-Dichloroethene-d2	4.015	65	67385	21.416	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	85.680%		
21) 2-Butanone-d5	7.075	46	65807	44.992	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	89.980%		
24) Chloroform-d	7.648	84	312491	24.307	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	97.240%		
26) 1,2-Dichloroethane-d4	8.301	65	164175	24.257	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	97.040%		
32) Benzene-d6	8.270	84	587018	27.512	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	110.040%		
36) 1,2-Dichloropropane-d6	9.270	67	186437	28.819	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	115.280%		
41) Toluene-d8	10.319	98	498477	26.180	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	104.720%		
43) trans-1,3-Dichloroprop...	10.575	79	56490	22.724	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	90.880%		
47) 2-Hexanone-d5	10.922	63	38440	45.026	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	90.060%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	103707	23.075	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	92.320%		
66) 1,2-Dichlorobenzene-d4	13.848	152	84297	23.959	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	95.840%		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	17898	2.646	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061424\
Data File : VW029149.D
Acq On : 14 Jun 2024 10:39
Operator : SY/MD
Sample : P2870-18
Misc : 5.84g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AB9

Quant Time: Jun 15 00:21:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060324SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jun 15 00:20:11 2024
Response via : Initial Calibration

